

IDWeek 2020: 907872

BACKGROUND

Identification of pathogens causing community acquired, health-care and ventilator associated pneumonia can be problematic. The BioFire® Pneumonia Plus (PNplus) Panel detects 15 bacteria (with semi-quantification), three atypical bacteria, eight viral classes and seven antibiotic resistance markers directly from sputum-like specimens (induced or expectorated sputum; endotracheal aspirates), and bronchoalveolar lavage (BAL)-like specimens with results in about one hour. This multinational study evaluated performance of BioFire PNplus Panel as compared to standard of care (SOC) testing.

Bacteria Semi – Quantitative Log Bins	Atypical Bacteria Qualitative	Resistance Markers Qualitative
<ul style="list-style-type: none"> <i>Acinetobacter calcoaceticus-baumannii</i> Complex⁺ <i>Enterobacter cloacae</i> complex⁺⁺ <i>Escherichia coli</i> <i>Klebsiella pneumoniae</i> group⁺⁺ <i>Klebsiella aerogenes</i> <i>Klebsiella oxytoca</i>⁺⁺ <i>Haemophilus influenzae</i> <i>Moraxella catarrhalis</i> <i>Proteus</i> spp.⁺⁺ <i>Pseudomonas aeruginosa</i>⁺ <i>Serratia marcescens</i>⁺⁺ <i>Staphylococcus aureus</i> <i>Streptococcus pneumoniae</i> <i>Streptococcus pyogenes</i> <i>Streptococcus agalactiae</i> 	<ul style="list-style-type: none"> <i>Legionella pneumophila</i> <i>Mycoplasma pneumoniae</i> <i>Chlamydia pneumoniae</i> <p>Viruses</p> <ul style="list-style-type: none"> Influenza A (H1, 09H1, H3) Influenza B Adenovirus Coronavirus (OC43, NL63, HKU1, 229E) Parainfluenza virus (1, 2, 3, 4) Respiratory syncytial virus (A, B) Human Rhinovirus/Enterovirus Human Metapneumovirus <p>OUS: includes MERS CoV</p>	<p>Methicillin resistance <i>mecA/mecC</i> and <i>MREJ</i></p> <p>ESBL CTX-M[*]</p> <p>Carbapenemases KPC⁺ NDM⁺ VIM⁺ IMP⁺ Oxa-48-like⁺</p> <p><small>* , +, ++ Denote when resistance marker would be reported based on pathogen detection Resistance markers cannot be directly linked to a specific pathogen</small></p>

MATERIALS AND METHODS

In total, 52 laboratories from 13 countries (Europe and Israel) compared BioFire PNplus Panel results to SOC results. SOC tests varied by site and included various combinations of Gram stain, culture, urinary antigen, molecular assays, and direct fluorescent antibody assays. 2,463 samples (1,237 sputum-like and 1,226 BAL-like) were evaluated. Values were applied to BioFire PNplus Panel semi-quantification results (bin results in copies/ml) (1=not detected (<10^{3.5}); 2=10⁴; 3=10⁵; 4=10⁶; and 5=≥10⁷) and compared to three algorithms of standardized SOC reporting (rare, few, moderate, numerous, 1+, 2+, 3+, 4+, quantitative culture etc.) to assess quantification correlation. All BioFire PNplus Panel and SOC results were considered true positive.

BACTERIA DETECTED BY BIOFIRE PNplus PANEL VS SOC (ALL SAMPLES)

Figure 1: 3,262 bacteria included in the BioFire PNplus were detected by at least one method. The BioFire PNplus Panel identified 3,125 (95.77%) bacteria compared to 1,861 (57.05%) for SOC (p<0.0001).




TABLE 1: PATHOGENS* DETECTED PER SAMPLE BY BIOFIRE PNplus PANEL VS SOC

Table 1: *Bacteria and viruses included in the BioFire PNplus Panel


	Total Positive N	1 Analyte N (%)	2 Analytes N (%)	3 Analytes N (%)	4 Analytes N (%)	5 Analytes N (%)	6+ Analytes N (%)
BioFire PNplus	1875	848 (45.22)	536 (28.59)	275 (14.67)	130 (6.93)	60 (3.20)	26 (1.39)
SOC	1380	904 (65.51)	370 (26.81)	85 (6.16)	15 (1.09)	4 (0.29)	2 (0.14)

Average number of pathogens* per positive sample: BioFire PNplus Panel = 1.99; SOC = 1.44

BACTERIA DETECTED BY BIOFIRE PNplus PANEL VS SOC BY SAMPLE TYPE

Figure 2: BAL-like samples: 1,358 bacteria included on BioFire PNplus Panel were detected by at least one method. The BioFire PNplus Panel identified 1,288 (94.35%) bacteria compared to 841 (61.93%) for SOC (p<0.0001).

Figure 3: Sputum-like samples: 1,904 bacteria included on BioFire PNplus Panel were detected by at least one method. The BioFire PNplus Panel identified 1,836 (96.43%) bacteria compared to 1020 (53.57%) for SOC (p<0.0001).



BAL-like Samples: Average number of bacteria/viruses* per positive sample: PNplus Panel = 1.8; SOC = 1.44

Sputum-like Samples: Average number of bacteria/viruses* per positive sample: PNplus Panel = 2.15; SOC = 1.48

*Bacteria and viruses included in BioFire PNplus Panel

TABLE 2: COMPARISON OF BIOFIRE PNplus PANEL VS SOC SEMI-QUANTIFICATION

Semi-quantification results for BioFire PNplus Panel and SOC bacteria were compared for 1,297 matched detections.

	Total N	PNplus = SOC Value (%)	PNplus 1 lg > SOC (%)	PNplus 2 lg > SOC (%)	PNplus 3 lg > SOC (%)	PNplus 4 lg > SOC (%)	SOC 1 lg > PNplus (%)	SOC 2 lg > PNplus (%)	SOC 3 lg > PNplus > FA (%)
All Samples	1297	329 (25.37)	477 (36.78)	282 (21.74)	123 (9.48)	20 (1.54%)	53 (4.09)	8 (0.62)	5 (0.39)
Total %			58.52		11.03				
Total %		25.37		69.55			5.91		
BAL-Like Samples	645	125 (19.38)	213 (33.02)	189 (29.30)	74 (11.47)	16 (2.48)	24 (3.72)	1 (0.16)	3 (0.49)
Total %			62.32		13.95				
Total %		19.38		76.28			4.34		
Sputum-like Samples	652	204 (31.29)	264 (40.49)	93 (14.26)	49 (7.52)	4 (0.61)	29 (4.45)	7 (1.07)	2 (0.31)
Total %			54.75		8.13				
Total %		31.92		62.88			5.83		

Lg = log₁₀; For the majority of samples (69.55%) BioFire PNplus Panel quantified bacteria higher than SOC methods. 25.4% of samples were reported with comparable quantification between BioFire PNplus and SOC and only 5.9% were quantified higher by SOC than BioFire PNplus Panel.

TABLE 3: DETECTION OF MSSA/MRSA BY BIOFIRE PNplus PANEL VS SOC

	Total + (%)	No FA+ (%)	No SOC (%)	No SOC+/FA+ No (%)	No SOC-/FA+ No (%)	No SOC+/FA- No (%)
<i>Staphylococcus aureus</i> (Total)	531	521 (98.12)	313 (58.95)	303 (57.06)	218 (41.05)	10 (1.88)
<i>Staphylococcus aureus</i> (MSSA)		412 ^a	269 ^b	234	178	11
<i>Staphylococcus aureus</i> (MRSA)		106 ^b	47 ^a	45	59	0

Overall, BioFire PNplus Panel detected more MSSA and MRSA than SOC. (+) = positive; (-) = negative
^aTwo samples reported to contain MRSA by SOC were reported to contain MSSA by BioFire PNplus Panel
^b24 samples reported to contain just MSSA by SOC were reported to contain MRSA by BioFire PNplus Panel

TABLE 4: DETECTION OF ATYPICAL PATHOGENS BY BIOFIRE PNplus PANEL VS SOC

Pathogen	PNplus	SOC ^A	Total Detected	% Detection PNplus	% Detection SOC	P Value
<i>Chlamydia pneumoniae</i>	7	4	7	100%	57.14	0.193
<i>Legionella pneumophila</i>	49	49	55	89.09	89.09	1.00
<i>Mycoplasma pneumoniae</i>	37	20	41	90.24	48.78	0.0001
Overall Detection Rate	93	73	103	90.29	70.87	<0.0001

^ALimited SOC site testing for typical pathogens: culture, molecular testing and/or urinary antigen

TABLE 5: DETECTION OF VIRAL PATHOGENS* BY BIOFIRE PNplus PANEL VS SOC

Pathogen	PNplus	SOC ^A	Total Detected	% Detection PNplus	% Detection SOC
Adenovirus	44	4	45	97.78	8.89
Coronavirus	51	6	52	98.08	11.54
Metapneumovirus	10	6	11	90.91	54.55
Rhinovirus/Enterovirus	365	60	368	99.18	16.30
Influenza A	44	25	45	97.78	55.56
Influenza B	2	2	2	100.00	100.00
Parainfluenza Virus	63	19	68	92.65	27.94
Respiratory Syncytial Virus	39	12	40	97.50	30.00
Overall Detection Rate	618	134	631	97.94	21.24

^AViruses on BioFire PNplus Panel only. Percent detection does not reflect direct comparison of testing by PNplus and SOC. SOC low detection rate was due to limited or no viral testing by clinical sites. There were 24 additional SOC viral detections reported for viruses not on BioFire PNplus Panel: 10 CMV; 5 EBV; 6 HSV; 2 HHV-6; 1 VZV

TABLE 6: CONCORDANCE BETWEEN BIOFIRE PNplus PANEL AND SOC (ALL SAMPLES)

Concordance	Number Total= 2463	Percentage (%)
Overall concordance	1209	49.10
Overall concordant positive	665	27.00
Concordant bacteria	601	24.40
Concordant bacteria and virus	25	1.02
Concordant virus	39	1.58
Concordant negative for any pathogen	412	16.73
Concordant negative for FilmArray pathogens	132	5.36
Partial concordance	651	26.43
No concordance	603	24.48

Overall concordance: PNplus Panel and SOC in agreement for positive and negative samples
Concordant positive: PNplus Panel and SOC in agreement for all pathogens detected
Concordant bacteria: PNplus Panel and SOC in agreement for all bacteria detected
Concordant bacteria and virus: PNplus Panel and SOC in agreement for all bacteria and viruses
Concordant virus: PNplus Panel and SOC in agreement for all viruses detected
Concordant negative for any pathogen: PNplus Panel and SOC both negative
Concordant negative for PNplus Panel pathogens: PNplus Panel and SOC negative for pathogens on PNplus Panel (other pathogens may have been detected by SOC)
Partial concordance: PNplus Panel and SOC in agreement for some pathogens detected
No concordance: PNplus Panel and SOC not in agreement

SUMMARY OF RESULTS

- Overall, and in both sample types, BioFire PNplus Panel detected more potential bacterial pathogens and at a higher concentration than SOC.
- Detection of viral pathogens was significantly higher for BioFire PNplus Panel as the majority of samples were not tested for these pathogens as part of their SOC testing. This highlights a missed opportunity to provide a more comprehensive and accurate diagnosis through a syndromic solution.
- The combined detection of the three atypical pathogens was higher for BioFire PNplus Panel than SOC. With the exception of *Legionella*, this difference was mainly due to a lack of SOC testing.
- The rapid and comprehensive detection of the cause of LRTI and antimicrobial resistance markers could lead to rapid initiation of targeted therapy.

Acknowledgement: BioFire Diagnostics, LLC sincerely thanks the EMEA Evaluation Collaborative for their participation in this study.
(Institutions and collaborators available by request)